



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2025/B/5809172
Dated/दिनांक : 28-01-2025

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	27-02-2025 16:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	27-02-2025 16:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	120 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Finance
Department Name/विभाग का नाम	Department Of Economic Affairs
Organisation Name/संगठन का नाम	Security Printing And Minting Corporation Of India Limited (spmci)
Office Name/कार्यालय का नाम	Janpath
क्रेता ईमेल/Buyer Email	buyer70.spmcils.ts@gembuyer.in
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	Endo-Bronchial Ultrasound System (EBUS)
GeMARPTS में खोजी गई स्ट्रिंग / Searched Strings used in GeMARPTS	Endo Bronchial Ultrasound System
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Endo Bronchial Ultrasound System (EBUS), Ultrasound Machine (V2), Endoscopic Ultrasound, Portable Ultrasound Machine (V2), Hand Held Ultrasound (V2), Echo Cardiography Scanning Machine (V2), High Definition Endovision system for Urology
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	• Endo Bronchial Ultrasound System (EBUS)
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	53 Lakh (s)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	53 Lakh (s)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes

Bid Details/बिड विवरण	
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No
Past Performance/विगत प्रदर्शन	40 %
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	Indusind bank
EMD Amount/ईएमडी राशि	263000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	Indusind bank
ePBG Percentage(%) /ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months)/ईपीबीजी की अपेक्षित अवधि (महीने).	38

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the

relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

CHIEF GENERAL MANAGER

P.B. NO: 10,IDA PHASE-II, CHERLAPALLY, HYDERABAD-500051, TELANGANA

(India Government Mint, Hyderabad)

UIN Number NCTGC2415P

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
5. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company

and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

6. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

7. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 40% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Endo-Bronchial Ultrasound System (EBUS) (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/कमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

Buyer Specification Document/क्रेता विशिष्टि दस्तावेज़	Download
--	--------------------------

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Hanumanthu Vasista Rao	500051,India Govt Mint, IDA, Phase-II Cherapally, (R.R district) Hyderabad , Telangana	1	60

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

THE SUPERINTENDENT,
GOVERNMENT GENERAL & CHEST HOSPITAL,
IRRUMNUMA
HYDERABAD
TELANGANA

2. Generic

Bidder financial standing: The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

3. Generic

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 7 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

4. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

5. Generic

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

6. Generic

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

7. Generic

IT equipment shall be IPv6 ready from day one.

8. Generic

Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

9. Generic

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

10. Generic

1. The Seller shall not assign the Contract in whole or part without obtaining the prior written consent of

buyer.

2. The Seller shall not sub-contract the Contract in whole or part to any entity without obtaining the prior written consent of buyer.

3. The Seller shall, notwithstanding the consent and assignment/sub-contract, remain jointly and severally liable and responsible to buyer together with the assignee/ sub-contractor, for and in respect of the due performance of the Contract and the Sellers obligations there under.

11. **Generic**

The seller is required to print logo as per buyer's requirement.

12. **Generic**

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

i) The Seller fails to comply with any material term of the Contract.

ii) The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.

iii) The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.

iv) The Seller becomes bankrupt or goes into liquidation.

v) The Seller makes a general assignment for the benefit of creditors.

vi) A receiver is appointed for any substantial property owned by the Seller.

vii) The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

13. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

14. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

15. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

16. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

17. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

18. **Warranty**

Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

19. Forms of EMD and PBG

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

India Government Mint, Hyderabad(A Unit of SPMCIL)
payable at
Hyderabad

. Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

20. Forms of EMD and PBG

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

India Government Mint, Hyderabad(A Unit of SPMCIL)
. The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

21. Forms of EMD and PBG

Bidders can also submit the EMD with Banker's Cheque in favour of

India Government Mint, Hyderabad(A Unit of SPMCIL)
payable at
Hyderabad

. Bidder has to upload scanned copy / proof of the BC along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

22. Forms of EMD and PBG

Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name

INDIA GOVERNMENT MINT, HYDERABAD (A UNIT OF SPMCIL) -COLLECTION A/C
Account No.
201003484303
IFSC Code
INDB0000004
Bank Name
INDUSIND BANK
Branch address
0004 - HYDERABAD

. Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

23. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

India Government Mint, Hyderabad(A Unit of SPMCIL)
payable at
Hyderabad

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

24. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name

INDIA GOVERNMENT MINT, HYDERABAD (A UNIT OF SPMCIL) -COLLECTION A/C

Account No.

201003484303

IFSC Code

INDB0000004

Bank Name

INDUSIND BANK

Branch address

0004 - HYDERABAD

. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

25. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

Technical specification for Endo-Bronchial Ultrasound System (EBUS):

I. Ultrasonic Broncho Video Scope (EBUS- TBNA):

1. Field of View at 120 degrees and above with camera on tip.
2. Depth of field : Approximately 3-50 mm or better
3. Direction of View:10-15 degree forward oblique or less.
4. Tip deflection : Up at least above 120 degrees Down at least 90 degrees
5. Distal end : 6.7 mm or less
6. Insertion tube outer diameter :6.3 mm or less
7. Instrument channel width : at least 2mm or more
8. Working length : At least 600mm
9. Acoustic frequency : 5-12 MHz or more
10. Scan angle -60 degrees or more
11. Convex linear Array Probe with balloon and direct contact option.
12. Scan direction parallel to insertion direction.
13. The scope should have camera on the tip of the Scope for crystal clear images.
14. Video Scope should preferably have separate channel for balloon inflation with saline which should be easily cleanable.
15. Scope should allow adequate visualization of the endoscopic view and exit of needle while performing EBUS- TBNA.
16. Hybrid scope should allow performance of bronchoscopy procedure like Endo-bronchial biopsy also.

II. Electronic Scan Digital Ultrasound Processor - 1 No

1. Compact & easily transportable unit with Ultrasound & Color Doppler function.
2. Ultrasonic processor should provide Colour Doppler, Power Doppler or effective confirmation of Blood flow, Pulse Wave, B Mode to see the volume & Velocity etc.
3. Scanning frequency :5/6/7.5/10/12 MHz or (any four range or more)
4. Ergonomic operation keyboard with user programmable calculation packing 7 Annotation, wow mark and pointer display.
5. To improve EBUS image resolution, system must be equipped with s

- ound speed correction technology.
6. System should be equipped with direct image storage during examination on Pen Drive.
 7. Should have Needle Guide Function to help precise targeting with minimum passes.
 8. Should have facility to measure distance between two points defined by 4 or more symbols.
 9. Touch screen, LED backlit keys and dedicated , user friendly key board.
 10. Memory: 150 frames or more
 11. Optimization function for auto setting of image by sensing the nature of tissue.
 12. Equipped hard disk drive with minimum 1-2GB to store still images.
 13. GAIN, Contrast, STC functions.
 14. Picture in Picture for the both ultrasound and endoscopic image simultaneously
 15. Should have dynamic tissue harmonic imaging and compound Harmonic.
 16. Should have tissue Electrography feature with Contrast harmonics.
 17. Video/image recording facility should be available.
 18. EBUS Bite guard-50 pieces should be provided.
 19. Compact & easily transportable one Cart system /unit , ease to move on trolley.

III. Video Processor & Light Source (Combined) - 1 No

1. Should be compatible with Analog, DVI-D, S Video x 1, Video output for a HDTV monitor should be available (Any HD outputs).
2. Should contain the electronics to operate Multi Optical Zoom for clear visibility of near & far objects.
3. Suitable for BLI/NBI/ISCAN-OE/FICE with real time optical chromo endoscopy system.
4. Equipped with high resolution HDTV imaging capacity.
5. Compact, lightweight (10-15 kg) and ergonomically designed.
6. Recording of both still & moving images.
7. Should be compatible with the Radial EBUS system and navigation system from the same manufacturer.
8. Portable Memory & USB Slot for image recording External memory USB (2GB)
9. Automatic IRIS control & automatic white balance
10. Electronic Zoom up to 1.5 X or better.
11. Equipped with 3 or more multi LED at least 5 years life of bulb/light.
12. Equipped with special filter for BLI&LCI.NBI & TXI/ISCAN-OE real time optical chromo endoscopy system.
13. Equipped with automatic light adjustment forced aircooling, regulated airfeeding pump and fan with low noise.

IV. Video Bronchoscope Slim - 1 No

1. Lighter and possess high resolution image quality.
2. Fully immersible in disinfectant solution.
3. Should have high resolution image with two light guide for better distribution of light.
4. Should have high definition image (Full HD Endoscopy) with tow light

t guide of better distribution of light.

5. Two or more no. of remote -control switches on control body.
6. Compatible with leakage testing device.
7. Scope should be compatible with NBI/FICE/ISCAN9OE) capacity.
8. Should be diode laser (810nm) compatible.
 - a. Field of view 120 Degree or More
 - b. Direction of View 0-degree, forward viewing
 - c. Depth of field 3 to 100 mm or better
 - d. Distal end outer diameter 4.9 mm or less
 - e. Insertion tube outer diameter 4.9 mm or less
 - f. Tip Bending range Up 180 deg & more. Down 130 deg & more
 - g. Working length 600 mm or more
 - h. Channel inner diameter 2.0 mm or more

V. Technical Specification for High Resolution Monitor - 1 No:

1. 1.27" Higher - definition medical grade monitor compatible with processor with full range of colors & inputs including wide viewing angle as needed for proper functioning.

VI. Computer, Trolley, Software etc for image recording and archiving Computer - 1 No

1. Intel i7 Processor, 8GB RAM, 1 TB hard disk with pre-installed operating system.
2. High speed USB slots (at least 2 in number)
3. 21" or more LCD color flat monitor with compatible mouse and keyboard.
4. UPS with at least 30- minute backup
5. Original Windows -10 software.
6. Color- Laser Printer
7. All essential accessories/ cords etc.

VII. Essential Accessories

1. EBUS aspiration needle set from the same company or US-FDA/CE approved, minimum 20 numbers ,22G-20 no. and 25/21G-10no.
2. EBUS balloon 40 no. (Minimum) preferably from the same company or USFDA/CE approved.
3. Mobile trolley with movable arm to move monitor and to mount the EBUS system and ultrasound.
4. Quoted rate for disposables, consumables be for next 5 years after warranty.
5. All others essentials/accessories required to make then machine function optimally.
6. Recording System for review &publication.
7. Should have compatible footswitch or other options for capturing images/videos during procedure.
8. Efficient and user-friendly HD software installed in computer for recording , retrieval of data, storage and easy copying of images/videos from computer to DVD/external hard disc.
9. Servo voltage stabilizer for the system.
10. Mouth Bite Guard-50
11. Cleaning Brush for working channels-10 each.
12. Balloon mounting forceps device-1no.

VIII. Other Terms and Conditions:

1. USFDA/CE/BIS approved.
2. The warranty of equipment should be 3 (Three) years from the date of installation.
3. Comprehensive Maintenance Contract (CMC) is NOT APPLICABLE.
4. Training of staff and personnel for proper usage and maintenance of the system.
5. Training of one Doctor at some reputed training center in India within six months of installation.

TERMS

1) The machines being supplied should carry the following credit slogan and SPMCIL LOGO prominently on it: "Machine Sponsored under CSR Funds for the Year 2024-25 by: India Government Mint, A Unit of SPMCIL, Hyderabad"

A. Eligibility Criteria:

The bidder should meet the following qualification criteria:

1. **Experience & Past Performance:** The Bidder (manufacturer or principal of authorized representative - hereinafter referred simply as 'The Bidder') shall be a manufacturer that has regularly for at least the last three years manufactured, supplied, Installed, commissioned Endo-Bronchial Ultrasound System (EBUS) with the same or higher specifications of at least one number in last five years ending on 31.03.2024. At least one number of the product offered for supply should be in successful operation for at least one year on the date of bid opening.
2. **Capability - Equipment & manufacturing Facilities:** The bidder must have an annual capacity to Supply, Install, commission Endo-Bronchial Ultrasound System (EBUS) with the same or higher specifications.
3. **Financial Standing:**
 - i. The average annual financial turnover of the bidder firm (manufacturer or principal of authorized dealer) during the last three years, ending on 31.03.2024 should be at Rs. 53 Lakhs as per the annual report (audited balance sheet and profit & loss account) of the relevant period duly authenticated by a Chartered Accountant/ Cost Accountant in India.
 - ii. The net worth of the firm should
 - i. not have eroded by more than 30% in the last three financial years ending 31.03.2024 and
 - ii. Net worth of the bidder should not be negative as on 31.03.2024.

Note: Participating bidder needs to submit all related documents along with the bid.

B. Undertaking :

1. Undertaking for acceptance of all the terms and conditions of the GeM Bid including Buyer Added Bid Specific Terms and conditions along with technical specification & scope of work without any deviation to be submitted duly sealed and signed.

C. The terms and conditions, guidelines of SPMCIL Procurement Manual Version 3.0 will be applicable to this bid at any stage to avoid any conflict at later stage. Kindly refer to GIT and GCC of SPMCIL Procurement Manual Version 3.0 for additional terms and condition as per the links given below:

- A. General Instruction to Tenderer(GIT):

<https://spmcil.com/uploaddocument/GIT3.0.pdf>

- B. General Conditions of Contract GCC):

<https://spmcil.com/uploaddocument/GCC3.0.pdf>

D. Special Instruction to Tenderers (SIT)

The following Special Instructions to Tenderers will apply for this purchase. These special instructions will modify/ substitute/ supplement the corresponding General Instructions to Tenderers (GIT). The corresponding GIT clause numbers have also been indicated in the text below. In case of any conflict between the provisions in the GIT and that in the SIT, the provisions contained in the SIT shall prevail.

1. **Signing and Sealing of Tender:** (GIT Clause No. 20.4 & 20.5) Tenderer shall sign all the pages of the tender document as token of the acceptance of all the tender conditions and upload through GeM portal.

E. Special Conditions of Contract (SCC)

The following Special Conditions of Contract (SCC) will apply for this purchase. The corresponding clauses of General Conditions of Contract (GCC) relating to the SCC stipulations have also been incorporated below. These Special Conditions will modify/substitute/supplement the corresponding (GCC) clauses.

Whenever there is any conflict between the provision in the GCC and that in the SCC, the provision contained in the SCC shall prevail.

1. **Terms of Delivery:** (GCC Clause No.10)

- 1.1. The Equipment Should be Supplied, Installed, Tested & Commissioned within 60 days from the date of issue of GeM Contract/NAC to "THE SUPERINTENDENT, GOVERNMENT GENERAL & CHEST HOSPITAL, IRRUMNUMA, HYDERABAD after verification at INDIA GOVERNMENT MINT, P.B. NO.10, HCL POST, IDA PHASE II, CHERLAPALLY, R.R. DIST., HYDERABAD - 500051.

2. Warranty: (GCC Clause No.16) -.

- i) The supplier should submit the warrantee certificate that the goods supplied under the contract are new, unused of the most recent or current models and incorporate all recent improvements in design and material unless provided otherwise in the contract.

- ii) The supplier further warrants that, all the goods supplied under this contract shall have no defect in design, materials or workmanship or from any act or omission of the supplier that may develop under normal use of the supplied goods in the conditions prevailing in the country of final destination.
- iii) This warranty shall remain valid for 36 months from the date of Installation. The purchaser shall promptly notify the supplier in writing of any claims arising out under this warranty.
- iv) Upon receipt of such notice, the supplier shall repair or replace the defective goods or parts thereof without cost to the purchaser immediately.
- v) If the supplier having been notified fails to take remedial measure or rectify the defect within the period specified, the purchaser may proceed to take such action as may be necessary, at the supplier's risk and expense without prejudice or any other rights, which the purchaser may have against the supplier under the contract.

3. Comprehensive Maintenance Contract (CMC):

- i) The Comprehensive Maintenance Contract (CMC) is NOT APPLICABLE.

4. Terms and Mode of Payments: (GCC Clause No.22.3, 22.4, 22.6)

- i) Payment shall be released within 30 days from the date of acceptance by Mint Authorities subject to online submission of Bills.

5. Training:

- i) Training of Staff and Personnel for proper usage and maintenance of the system.
- ii) Training of One doctor at some reputed training center in India within Six months of Installation.

F. Check list for the bidder to submit the documents in Technical bid:

1	Experience & Past Performance: Documents in support
2	Capability - Equipment & Manufacturing Facilities: Documents in support
3	Financial Standing: Documents in support
4	Undertakings
4.1	Undertaking to all Technical Specifications & Scope of work without any deviation DULY STAMPED AND SIGNED
4.2	Undertaking to all terms and conditions as set forth in this GeM bid including Buyer Added Bid specific ATC Clauses.
5	Acceptance of GIT & GCC as per the links provided in the Buyer added bid specific ATC Clauses
6	Any other document as per tender

–
NOTE: Any dispute arising out of this contract shall be subject to Hyderabad Jurisdiction only.

26. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---